

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031619\
 Data File : VX008146.D
 Acq On : 16 Mar 2019 17:34
 Operator : JC/SP
 Sample : K1685-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 AU-05-031419-C

Quant Time: Mar 18 02:16:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X031419W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 15 03:56:35 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	315443	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	429358	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	365366	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	164359	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	141401	44.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.88%	
35) Dibromofluoromethane	5.50	113	130417	47.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.80%	
50) Toluene-d8	8.71	98	494447	47.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.50%	
62) 4-Bromofluorobenzene	11.13	95	159917	43.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.22%	
Target Compounds						
16) Acetone	2.45	43	12485	8.311	ug/l	90
18) Methyl Acetate	2.78	43	6091	1.876	ug/l #	90
20) Methylene Chloride	2.86	84	105492	39.319	ug/l	90
43) Isopropyl Acetate	6.46	43	7929	1.297	ug/l	98
95) Naphthalene	13.83	128	74445	7.367	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031619\
Data File : VX008146.D
Acq On : 16 Mar 2019 17:34
Operator : JC/SP
Sample : K1685-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
AU-05-031419-C

Quant Time: Mar 18 02:16:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X031419W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 15 03:56:35 2019
Response via : Initial Calibration

