

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031619\
 Data File : VX008147.D
 Acq On : 16 Mar 2019 18:01
 Operator : JC/SP
 Sample : K1685-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 AU-05-031419-D

Quant Time: Mar 18 02:23:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X031419W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 15 03:56:35 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	319450	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	429482	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	368290	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	157447	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	140261	43.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.06%	
35) Dibromofluoromethane	5.50	113	130664	47.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.96%	
50) Toluene-d8	8.71	98	497494	48.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.06%	
62) 4-Bromofluorobenzene	11.13	95	156270	42.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.22%	
Target Compounds						
16) Acetone	2.45	43	8823	5.800	ug/l	96
20) Methylene Chloride	2.84	84	25991439	9565.995	ug/l #	74
43) Isopropyl Acetate	6.46	43	9738	1.592	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031619\
Data File : VX008147.D
Acq On : 16 Mar 2019 18:01
Operator : JC/SP
Sample : K1685-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
AU-05-031419-D

Quant Time: Mar 18 02:23:44 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X031419W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 15 03:56:35 2019
Response via : Initial Calibration

