

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031619\
 Data File : VX008148.D
 Acq On : 16 Mar 2019 18:27
 Operator : JC/SP
 Sample : K1905-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RBR-200052

Quant Time: Mar 18 02:35:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X031419W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 15 03:56:35 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	315237	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	428180	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	372793	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	172892	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	139611	43.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.80%	
35) Dibromofluoromethane	5.51	113	130293	47.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.98%	
50) Toluene-d8	8.71	98	498685	48.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.58%	
62) 4-Bromofluorobenzene	11.13	95	163660	44.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.48%	
Target Compounds						
16) Acetone	2.45	43	10815	7.204	ug/l	91
18) Methyl Acetate	2.78	43	4688	1.445	ug/l	100
20) Methylene Chloride	2.86	84	66186	24.685	ug/l	89
43) Isopropyl Acetate	6.46	43	8341	1.368	ug/l #	96
95) Naphthalene	13.83	128	11246	1.058	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031619\
Data File : VX008148.D
Acq On : 16 Mar 2019 18:27
Operator : JC/SP
Sample : K1905-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RBR-200052

Quant Time: Mar 18 02:35:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X031419W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 15 03:56:35 2019
Response via : Initial Calibration

