

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031619\
 Data File : VX008153.D
 Acq On : 16 Mar 2019 20:40
 Operator : JC/SP
 Sample : K1960-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T3-MW-2-9.0-031419

Manual Integrations
 APPROVED

MMDadoda
 3/18/2019 10:37:05 AM

Quant Time: Mar 18 03:26:59 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X031419W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 15 03:56:35 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	332773	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	444389	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	378346	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	175431	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	141487	42.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.30%	
35) Dibromofluoromethane	5.51	113	132064	46.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.72%	
50) Toluene-d8	8.71	98	514797	48.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.06%	
62) 4-Bromofluorobenzene	11.13	95	173827	45.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.54%	
Target Compounds						
					Qvalue	
16) Acetone	2.45	43	154521	97.504	ug/l	95
18) Methyl Acetate	2.78	43	3643	1.064	ug/l #	87
20) Methylene Chloride	2.86	84	2954	1.044	ug/l #	83
31) Cyclohexane	5.58	56	31344	6.530	ug/l #	92
39) Methylcyclohexane	7.46	83	97875	20.945	ug/l	100
40) Benzene	6.15	78	3739	0.334	ug/l #	86
52) Toluene	8.79	92	3506	0.492	ug/l	93
67) Ethyl Benzene	10.25	91	10936	0.880	ug/l	96
73) Isopropylbenzene	11.02	105	433866	38.614	ug/l	100
78) n-propylbenzene	11.36	91	1845293	149.691	ug/l	99
83) tert-Butylbenzene	11.77	119	25794	2.780	ug/l	92
85) sec-Butylbenzene	11.94	105	169345	15.319	ug/l	96
89) n-Butylbenzene	12.38	91	146885m	17.558	ug/l	
95) Naphthalene	13.83	128	25879	2.399	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031619\
Data File : VX008153.D
Acq On : 16 Mar 2019 20:40
Operator : JC/SP
Sample : K1960-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T3-MW-2-9.0-031419

Manual Integrations
APPROVED

MMDadoda
3/18/2019 10:37:05 AM

Quant Time: Mar 18 03:26:59 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X031419W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 15 03:56:35 2019
Response via : Initial Calibration

