

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031619\
 Data File : VX008156.D
 Acq On : 16 Mar 2019 22:00
 Operator : JC/SP
 Sample : K1908-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 101-102

Quant Time: Mar 18 03:38:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X031419W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 15 03:56:35 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	318091	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	429583	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	372788	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	172686	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	137717	42.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.84%	
35) Dibromofluoromethane	5.51	113	128782	47.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.54%	
50) Toluene-d8	8.71	98	498177	48.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.16%	
62) 4-Bromofluorobenzene	11.13	95	166196	44.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.56%	
Target Compounds						
16) Acetone	2.45	43	8675	5.727	ug/l	96
18) Methyl Acetate	2.78	43	5307	1.621	ug/l #	85
20) Methylene Chloride	2.86	84	5960	2.203	ug/l	94
43) Isopropyl Acetate	6.46	43	14457	2.363	ug/l	98
95) Naphthalene	13.83	128	13751	1.295	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031619\
Data File : VX008156.D
Acq On : 16 Mar 2019 22:00
Operator : JC/SP
Sample : K1908-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
101-102

Quant Time: Mar 18 03:38:47 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X031419W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 15 03:56:35 2019
Response via : Initial Calibration

