

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031620\
 Data File : VX015287.D
 Acq On : 16 Mar 2020 12:34
 Operator : JC/SP
 Sample : L1903-17DL 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT101D1-20200310DL

Quant Time: Mar 17 08:07:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 04 14:19:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	280242	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	440541	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	424457	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	232403	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	155709	46.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.66%	
35) Dibromofluoromethane	5.48	113	134109	48.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.50%	
50) Toluene-d8	8.71	98	531660	50.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.58%	
62) 4-Bromofluorobenzene	11.13	95	208016	55.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.32%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.37	101	10239	3.880	ug/l	97
12) 1,1-Dichloroethene	2.36	96	4040	1.483	ug/l #	81
44) Trichloroethene	7.20	130	160974	47.838	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031620\
Data File : VX015287.D
Acq On : 16 Mar 2020 12:34
Operator : JC/SP
Sample : L1903-17DL 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT101D1-20200310DL

Quant Time: Mar 17 08:07:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 04 14:19:10 2020
Response via : Initial Calibration

