

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031620\  
 Data File : VX015289.D  
 Acq On : 16 Mar 2020 13:19  
 Operator : JC/SP  
 Sample : L1903-09DL 10X  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 9 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 RE120D1-20200309DL

Quant Time: Mar 17 08:13:51 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X030420W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Mar 04 14:19:10 2020  
 Response via : Initial Calibration

| Internal Standards            | R.T.  | QIon | Response | Conc   | Units   | Dev(Min) |
|-------------------------------|-------|------|----------|--------|---------|----------|
| 1) Pentafluorobenzene         | 5.65  | 168  | 273143   | 50.00  | ug/l    | 0.00     |
| 34) 1,4-Difluorobenzene       | 6.85  | 114  | 435102   | 50.00  | ug/l    | 0.00     |
| 63) Chlorobenzene-d5          | 10.11 | 117  | 431038   | 50.00  | ug/l    | 0.00     |
| 72) 1,4-Dichlorobenzene-d4    | 12.07 | 152  | 229117   | 50.00  | ug/l    | 0.00     |
| System Monitoring Compounds   |       |      |          |        |         |          |
| 33) 1,2-Dichloroethane-d4     | 6.05  | 65   | 154203   | 47.59  | ug/l    | 0.00     |
| Spiked Amount 50.000          |       |      | Recovery | =      | 95.18%  |          |
| 35) Dibromofluoromethane      | 5.48  | 113  | 131903   | 48.05  | ug/l    | 0.00     |
| Spiked Amount 50.000          |       |      | Recovery | =      | 96.10%  |          |
| 50) Toluene-d8                | 8.71  | 98   | 534924   | 51.23  | ug/l    | 0.00     |
| Spiked Amount 50.000          |       |      | Recovery | =      | 102.46% |          |
| 62) 4-Bromofluorobenzene      | 11.13 | 95   | 207347   | 55.67  | ug/l    | 0.00     |
| Spiked Amount 50.000          |       |      | Recovery | =      | 111.34% |          |
| Target Compounds              |       |      |          |        |         |          |
| 9) 1,1,2-Trichlorotrifluoroet | 2.36  | 101  | 5339     | 2.076  | ug/l    | 91       |
| 12) 1,1-Dichloroethene        | 2.36  | 96   | 3322     | 1.252  | ug/l    | 93       |
| 44) Trichloroethene           | 7.20  | 130  | 302331   | 90.969 | ug/l    | 95       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031620\  
Data File : VX015289.D  
Acq On : 16 Mar 2020 13:19  
Operator : JC/SP  
Sample : L1903-09DL 10X  
Misc : 5.0mL/MSVOA X/WATER  
ALS Vial : 9 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
RE120D1-20200309DL

Quant Time: Mar 17 08:13:51 2020  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_X\METHOD\82X030420W.M  
Quant Title : SW846 8260  
QLast Update : Wed Mar 04 14:19:10 2020  
Response via : Initial Calibration

