

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031620\
 Data File : VX015290.D
 Acq On : 16 Mar 2020 13:42
 Operator : JC/SP
 Sample : L1903-10DL 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE120D2-20200309DL

Quant Time: Mar 17 08:15:48 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 04 14:19:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	330932	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	524359	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	514899	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	263208	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	187557	47.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.54%	
35) Dibromofluoromethane	5.49	113	162747	49.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.40%	
50) Toluene-d8	8.71	98	636488	50.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.16%	
62) 4-Bromofluorobenzene	11.13	95	242258	53.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.96%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.38	101	6940	2.227	ug/l	94
44) Trichloroethene	7.21	130	295019	73.658	ug/l	99
64) Tetrachloroethene	9.33	164	2767	0.650	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031620\
Data File : VX015290.D
Acq On : 16 Mar 2020 13:42
Operator : JC/SP
Sample : L1903-10DL 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE120D2-20200309DL

Quant Time: Mar 17 08:15:48 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 04 14:19:10 2020
Response via : Initial Calibration

