

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031620\
 Data File : VX015299.D
 Acq On : 16 Mar 2020 17:05
 Operator : JC/SP
 Sample : L1945-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE104D2-20200312

Quant Time: Mar 17 08:37:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 04 14:19:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	318572	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	520164	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	506206	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	261287	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	182144	48.19	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.38%	
35) Dibromofluoromethane	5.49	113	158627	48.34	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.68%	
50) Toluene-d8	8.71	98	623885	49.98	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.96%	
62) 4-Bromofluorobenzene	11.13	95	238601	53.59	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.18%	
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.37	96	2001	0.646	ug/l #	63
21) trans-1,2-Dichloroethene	3.16	96	6063	1.834	ug/l	84
24) 1,1-Dichloroethane	3.69	63	8200	1.374	ug/l #	89
27) cis-1,2-Dichloroethene	4.59	96	130658	35.232	ug/l	90
30) Chloroform	5.20	83	22058	3.792	ug/l	93
42) 1,2-Dichloroethane	6.19	62	8360	1.706	ug/l	91
44) Trichloroethene	7.21	130	878486	221.103	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031620\
Data File : VX015299.D
Acq On : 16 Mar 2020 17:05
Operator : JC/SP
Sample : L1945-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE104D2-20200312

Quant Time: Mar 17 08:37:17 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 04 14:19:10 2020
Response via : Initial Calibration

