

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031620\
 Data File : VX015303.D
 Acq On : 16 Mar 2020 18:58
 Operator : JC/SP
 Sample : L1945-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP05-20200312

Quant Time: Mar 17 08:53:50 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 04 14:19:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	302032	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	496538	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	482635	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	249459	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	177733	49.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.20%	
35) Dibromofluoromethane	5.49	113	153969	49.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.30%	
50) Toluene-d8	8.71	98	613279	51.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.94%	
62) 4-Bromofluorobenzene	11.13	95	223031	52.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.96%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.59	96	2282	0.649	ug/l	86
44) Trichloroethene	7.21	130	174819	46.093	ug/l	97
64) Tetrachloroethene	9.33	164	11083	2.777	ug/l	97
95) Naphthalene	13.83	128	5397	1.737	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031620\
Data File : VX015303.D
Acq On : 16 Mar 2020 18:58
Operator : JC/SP
Sample : L1945-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP05-20200312

Quant Time: Mar 17 08:53:50 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 04 14:19:10 2020
Response via : Initial Calibration

