

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031621\
 Data File : VX021246.D
 Acq On : 16 Mar 2021 15:46
 Operator : JC/MD
 Sample : VX0316WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0316WBL01

Quant Time: Mar 17 02:39:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 16 14:00:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	82307	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.829	114	134312	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	112958	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.065	152	46786	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	62950	46.63	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.26%
35) Dibromofluoromethane	5.464	113	44476	48.66	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.32%
50) Toluene-d8	8.694	98	164045	49.47	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.94%
62) 4-Bromofluorobenzene	11.118	95	57768	44.86	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	89.72%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031621\
Data File : VX021246.D
Acq On : 16 Mar 2021 15:46
Operator : JC/MD
Sample : VX0316WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0316WBL01

Quant Time: Mar 17 02:39:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 16 14:00:21 2021
Response via : Initial Calibration

