

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031621\
 Data File : VX021255.D
 Acq On : 16 Mar 2021 19:30
 Operator : JC/MD
 Sample : M1690-01 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MC-WC

Quant Time: Mar 17 02:43:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 16 14:00:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	77498	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	123835	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	105745	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	43172	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.035	65	58600	46.10	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	92.20%		
35) Dibromofluoromethane	5.464	113	41627	49.40	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	98.80%		
50) Toluene-d8	8.694	98	153376	50.17	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	100.34%		
62) 4-Bromofluorobenzene	11.118	95	54502	45.90	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	91.80%		
Target Compounds						
16) Acetone	2.429	43	2147	4.59	ug/l	100
84) 1,2,4-Trimethylbenzene	11.789	105	1124	0.41	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031621\
Data File : VX021255.D
Acq On : 16 Mar 2021 19:30
Operator : JC/MD
Sample : M1690-01 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MC-WC

Quant Time: Mar 17 02:43:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 16 14:00:21 2021
Response via : Initial Calibration

