

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031622\
 Data File : VX027471.D
 Acq On : 16 Mar 2022 15:16
 Operator : JC/MD
 Sample : N1950-02RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 31340RE

Quant Time: Mar 17 00:51:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 15:54:34 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	364211	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	588412	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	494482	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	191058	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	228171	53.722	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 107.440%	
35) Dibromofluoromethane	5.391	113	197984	52.145	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 104.300%	
50) Toluene-d8	8.653	98	681150	47.520	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 95.040%	
62) 4-Bromofluorobenzene	11.085	95	223093	40.836	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 81.680%	
Target Compounds						
						Qvalue
11) Tert butyl alcohol	2.965	59	29457	35.378	ug/l	98
16) Acetone	2.379	43	346617	210.111	ug/l	98
25) 2-Butanone	4.562	43	83728	28.619	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031622\
Data File : VX027471.D
Acq On : 16 Mar 2022 15:16
Operator : JC/MD
Sample : N1950-02RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
31340RE

Quant Time: Mar 17 00:51:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 15:54:34 2022
Response via : Initial Calibration

