

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031622\
 Data File : VX027475.D
 Acq On : 16 Mar 2022 16:50
 Operator : JC/MD
 Sample : N1977-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-20220315

Quant Time: Mar 17 00:52:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 15:54:34 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	355869	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	579467	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	534361	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	251861	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	216741	52.227	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	104.460%
35) Dibromofluoromethane	5.391	113	188907	50.522	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.040%
50) Toluene-d8	8.653	98	682289	48.334	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.660%
62) 4-Bromofluorobenzene	11.085	95	257350	47.834	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.660%
Target Compounds						
16) Acetone	2.380	43	8019	2.122	ug/l	94
20) Methylene Chloride	2.788	84	1529	0.377	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031622\
Data File : VX027475.D
Acq On : 16 Mar 2022 16:50
Operator : JC/MD
Sample : N1977-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-20220315

Quant Time: Mar 17 00:52:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 15:54:34 2022
Response via : Initial Calibration

