

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031622\
 Data File : VX027477.D
 Acq On : 16 Mar 2022 17:47
 Operator : JC/MD
 Sample : N1977-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-20220315

Quant Time: Mar 17 00:52:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 15:54:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	352408	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	570774	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	576974	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	242909	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	227108	55.263	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 110.520%	
35) Dibromofluoromethane	5.385	113	195423	53.061	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 106.120%	
50) Toluene-d8	8.647	98	704368	50.658	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 101.320%	
62) 4-Bromofluorobenzene	11.085	95	275088	51.910	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 103.820%	
Target Compounds						
16) Acetone	2.380	43	7490	1.835	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031622\
Data File : VX027477.D
Acq On : 16 Mar 2022 17:47
Operator : JC/MD
Sample : N1977-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-20220315

Quant Time: Mar 17 00:52:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 15:54:34 2022
Response via : Initial Calibration

