

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031622\
 Data File : VX027478.D
 Acq On : 16 Mar 2022 18:11
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 17 00:52:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 15:54:34 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	81	0.00
2 T	Dichlorodifluoromethane	0.513	0.478	6.8	76	0.00
3 P	Chloromethane	0.539	0.412	23.6	67	0.00
4 C	Vinyl Chloride	0.554	0.478	13.7#	72	0.00
5 T	Bromomethane	0.243	0.226	7.0	81	0.00
6 T	Chloroethane	0.346	0.304	12.1	81	0.00
7 T	Trichlorofluoromethane	0.771	0.801	-3.9	89	0.00
8 T	Diethyl Ether	0.301	0.306	-1.7	86	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.504	0.473	6.2	81	0.00
10 T	Methyl Iodide	0.712	0.612	14.0	69	0.00
11 T	Tert butyl alcohol	0.114	0.121	-6.1	88	0.00
12 CM	1,1-Dichloroethene	0.485	0.451	7.0#	77	0.00
13 T	Acrolein	0.080	0.081	-1.3	91	0.00
14 T	Allyl chloride	0.760	0.715	5.9	78	0.00
15 T	Acrylonitrile	0.301	0.277	8.0	77	0.00
16 T	Acetone	0.256	0.246	3.9	86	0.00
17 T	Carbon Disulfide	1.317	1.001	24.0	66	0.00
18 T	Methyl Acetate	0.641	0.617	3.7	81	0.00
19 T	Methyl tert-butyl Ether	1.606	1.648	-2.6	86	0.00
20 T	Methylene Chloride	0.570	0.507	11.1	76	0.00
21 T	trans-1,2-Dichloroethene	0.539	0.473	12.2	76	0.00
22 T	Diisopropyl ether	1.520	1.484	2.4	80	0.00
23 T	Vinyl Acetate	1.272	1.308	-2.8	83	0.00
24 P	1,1-Dichloroethane	0.923	0.889	3.7	80	0.00
25 T	2-Butanone	0.402	0.390	3.0	82	0.00
26 T	2,2-Dichloropropane	0.662	0.682	-3.0	85	0.00
27 T	cis-1,2-Dichloroethene	0.625	0.581	7.0	79	0.00
28 T	Bromochloromethane	0.384	0.341	11.2	75	0.00
29 T	Tetrahydrofuran	0.266	0.252	5.3	79	0.00
30 C	Chloroform	0.974	0.983	-0.9#	85	0.00
31 T	Cyclohexane	0.816	0.724	11.3	73	0.00
32 T	1,1,1-Trichloroethane	0.833	0.884	-6.1	88	0.00
33 S	1,2-Dichloroethane-d4	0.583	0.689	-18.2	99	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	80	0.00
35 S	Dibromofluoromethane	0.323	0.367	-13.6	92	0.00
36 T	1,1-Dichloropropene	0.448	0.423	5.6	78	0.00
37 T	Ethyl Acetate	0.472	0.470	0.4	82	0.00
38 T	Carbon Tetrachloride	0.463	0.496	-7.1	87	0.00
39 T	Methylcyclohexane	0.566	0.563	0.5	82	0.00
40 TM	Benzene	1.364	1.252	8.2	77	0.00
41 T	Methacrylonitrile	0.255	0.258	-1.2	81	0.00
42 TM	1,2-Dichloroethane	0.446	0.495	-11.0	90	0.00
43 T	Isopropyl Acetate	0.712	0.753	-5.8	86	0.00
44 TM	Trichloroethene	0.375	0.404	-7.7	90	0.00
45 C	1,2-Dichloropropane	0.342	0.352	-2.9#	84	0.00
46 T	Dibromomethane	0.243	0.261	-7.4	88	0.00
47 T	Bromodichloromethane	0.449	0.498	-10.9	90	0.00
48 T	Methyl methacrylate	0.354	0.382	-7.9	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031622\
 Data File : VX027478.D
 Acq On : 16 Mar 2022 18:11
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 17 00:52:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 15:54:34 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.009	0.0	85	0.00
50 S	Toluene-d8	1.218	1.387	-13.9	93	0.00
51 T	4-Methyl-2-Pentanone	0.477	0.502	-5.2	84	0.00
52 CM	Toluene	0.870	0.877	-0.8#	82	0.00
53 T	t-1,3-Dichloropropene	0.461	0.504	-9.3	85	0.00
54 T	cis-1,3-Dichloropropene	0.517	0.571	-10.4	87	0.00
55 T	1,1,2-Trichloroethane	0.363	0.364	-0.3	82	0.00
56 T	Ethyl methacrylate	0.535	0.553	-3.4	82	0.00
57 T	1,3-Dichloropropane	0.586	0.576	1.7	81	0.00
58 T	2-Chloroethyl Vinyl ether	0.270	0.317	-17.4	93	0.00
59 T	2-Hexanone	0.368	0.379	-3.0	81	0.00
60 T	Dibromochloromethane	0.370	0.402	-8.6	86	0.00
61 T	1,2-Dibromoethane	0.375	0.384	-2.4	83	0.00
62 S	4-Bromofluorobenzene	0.464	0.500	-7.8	88	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	79	0.00
64 T	Tetrachloroethene	0.350	0.367	-4.9	88	0.00
65 PM	Chlorobenzene	1.052	1.023	2.8	80	0.00
66 T	1,1,1,2-Tetrachloroethane	0.368	0.396	-7.6	87	0.00
67 C	Ethyl Benzene	1.814	1.797	0.9#	80	0.00
68 T	m/p-Xylenes	0.712	0.699	1.8	79	0.00
69 T	o-Xylene	0.703	0.697	0.9	80	0.00
70 T	Styrene	1.166	1.161	0.4	79	0.00
71 P	Bromoform	0.300	0.327	-9.0	85	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	78	0.00
73 T	Isopropylbenzene	3.657	3.679	-0.6	80	0.00
74 T	N-amyl acetate	1.337	1.437	-7.5	83	0.00
75 P	1,1,2,2-Tetrachloroethane	1.257	1.196	4.9	78	0.00
76 T	1,2,3-Trichloropropane	1.097	1.358	-23.8	98	0.00
77 T	Bromobenzene	0.923	0.923	0.0	80	0.00
78 T	n-propylbenzene	4.113	4.066	1.1	77	0.00
79 T	2-Chlorotoluene	2.571	2.520	2.0	79	0.00
80 T	1,3,5-Trimethylbenzene	3.135	3.141	-0.2	80	0.00
81 T	trans-1,4-Dichloro-2-butene	0.339	0.349	-2.9	79	0.00
82 T	4-Chlorotoluene	2.908	2.859	1.7	78	0.00
83 T	tert-Butylbenzene	3.134	3.158	-0.8	80	0.00
84 T	1,2,4-Trimethylbenzene	3.109	3.140	-1.0	79	0.00
85 T	sec-Butylbenzene	3.780	3.806	-0.7	79	0.00
86 T	p-Isopropyltoluene	3.200	3.250	-1.6	79	0.00
87 T	1,3-Dichlorobenzene	1.734	1.716	1.0	80	0.00
88 T	1,4-Dichlorobenzene	1.755	1.701	3.1	80	0.00
89 T	n-Butylbenzene	2.680	2.695	-0.6	77	0.00
90 T	Hexachloroethane	0.499	0.526	-5.4	83	0.00
91 T	1,2-Dichlorobenzene	1.709	1.689	1.2	81	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.263	0.283	-7.6	85	0.00
93 T	1,2,4-Trichlorobenzene	1.122	1.078	3.9	77	0.00
94 T	Hexachlorobutadiene	0.492	0.466	5.3	77	0.00
95 T	Naphthalene	3.767	3.732	0.9	76	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031622\
Data File : VX027478.D
Acq On : 16 Mar 2022 18:11
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Mar 17 00:52:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 15:54:34 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.121	1.087	3.0	77	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6