

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031623\
 Data File : VX034537.D
 Acq On : 16 Mar 2023 11:52
 Operator : JC/MD
 Sample : 01901-01
 Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TR-04-031323

Quant Time: Mar 17 06:22:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	243557	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	412521	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	361940	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	172064	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	169766	47.912	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.820%
35) Dibromofluoromethane	5.385	113	127801	46.923	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.840%
50) Toluene-d8	8.646	98	499005	49.078	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.160%
62) 4-Bromofluorobenzene	11.085	95	181697	44.242	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031623\
Data File : VX034537.D
Acq On : 16 Mar 2023 11:52
Operator : JC/MD
Sample : 01901-01
Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TR-04-031323

Quant Time: Mar 17 06:22:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

