

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031623\
 Data File : VX034538.D
 Acq On : 16 Mar 2023 12:16
 Operator : JC/MD
 Sample : PB151426TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB151426TB

Quant Time: Mar 17 06:23:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	264067	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	456208	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	457560	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	193173	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	184499	48.026	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.060%
35) Dibromofluoromethane	5.385	113	141705	47.045	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.100%
50) Toluene-d8	8.647	98	557953	49.621	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.240%
62) 4-Bromofluorobenzene	11.079	95	253947	55.913	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	111.820%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.940	59	17965	22.456	ug/l #	79
16) Acetone	2.373	43	79139	43.733	ug/l	99
20) Methylene Chloride	2.788	84	10889	3.100	ug/l	97
37) Ethyl Acetate	4.708	43	68685	13.590	ug/l	99
43) Isopropyl Acetate	6.336	43	283894	33.157	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031623\
Data File : VX034538.D
Acq On : 16 Mar 2023 12:16
Operator : JC/MD
Sample : PB151426TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB151426TB

Quant Time: Mar 17 06:23:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

