

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031623\
 Data File : VX034540.D
 Acq On : 16 Mar 2023 13:03
 Operator : JC/MD
 Sample : PB151426ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB151426ZHE#02

Quant Time: Mar 17 06:23:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	253218	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	434080	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	390565	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	165075	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	173766	47.170	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.340%
35) Dibromofluoromethane	5.385	113	136919	47.774	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.540%
50) Toluene-d8	8.647	98	529598	49.500	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.000%
62) 4-Bromofluorobenzene	11.079	95	201426	46.610	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.220%
Target Compounds						
					Qvalue	
16) Acetone	2.373	43	50969	29.373	ug/l	98
20) Methylene Chloride	2.782	84	14349	4.261	ug/l	93
37) Ethyl Acetate	4.715	43	70532	14.667	ug/l	99
43) Isopropyl Acetate	6.336	43	291721	35.808	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031623\
Data File : VX034540.D
Acq On : 16 Mar 2023 13:03
Operator : JC/MD
Sample : PB151426ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB151426ZHE#02

Quant Time: Mar 17 06:23:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

