

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031623\
 Data File : VX034541.D
 Acq On : 16 Mar 2023 13:26
 Operator : JC/MD
 Sample : PB151426ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB151426ZHE#03

Quant Time: Mar 17 06:24:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	254934	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	440451	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	431859	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	170073	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	177201	47.779	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.560%
35) Dibromofluoromethane	5.385	113	136327	46.879	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.760%
50) Toluene-d8	8.647	98	537289	49.492	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.980%
62) 4-Bromofluorobenzene	11.079	95	231868	52.878	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.760%
Target Compounds						
					Qvalue	
16) Acetone	2.373	43	39738	22.746	ug/l	95
18) Methyl Acetate	2.703	43	6431	1.127	ug/l	94
20) Methylene Chloride	2.788	84	14057	4.146	ug/l	97
37) Ethyl Acetate	4.708	43	73792	15.122	ug/l	99
43) Isopropyl Acetate	6.336	43	300824	36.391	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031623\
Data File : VX034541.D
Acq On : 16 Mar 2023 13:26
Operator : JC/MD
Sample : PB151426ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB151426ZHE#03

Quant Time: Mar 17 06:24:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

