

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031623\
 Data File : VX034542.D
 Acq On : 16 Mar 2023 13:50
 Operator : JC/MD
 Sample : PB151426ZHE#04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB151426ZHE#04

Quant Time: Mar 17 06:24:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	245982	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	426267	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	383837	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	182044	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	174211	48.682	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.360%
35) Dibromofluoromethane	5.385	113	131318	46.659	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.320%
50) Toluene-d8	8.647	98	522970	49.776	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.560%
62) 4-Bromofluorobenzene	11.079	95	230781	54.381	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.760%
Target Compounds						
					Qvalue	
16) Acetone	2.374	43	40330	23.925	ug/l	96
18) Methyl Acetate	2.703	43	6479	1.176	ug/l	92
20) Methylene Chloride	2.788	84	13998	4.279	ug/l	97
37) Ethyl Acetate	4.715	43	73262	15.513	ug/l	98
43) Isopropyl Acetate	6.336	43	303167	37.895	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031623\
Data File : VX034542.D
Acq On : 16 Mar 2023 13:50
Operator : JC/MD
Sample : PB151426ZHE#04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB151426ZHE#04

Quant Time: Mar 17 06:24:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

