

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031623\
 Data File : VX034543.D
 Acq On : 16 Mar 2023 14:14
 Operator : JC/MD
 Sample : PB151426ZHE#05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB151426ZHE#05

Quant Time: Mar 17 06:25:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	257524	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	444991	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	410656	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	185722	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	180207	48.101	ug/l	0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.200%
35) Dibromofluoromethane	5.385	113	139173	47.369	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.740%
50) Toluene-d8	8.647	98	547133	49.885	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.760%
62) 4-Bromofluorobenzene	11.079	95	219799	49.614	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.220%
Target Compounds						
					Qvalue	
16) Acetone	2.374	43	39351	22.298	ug/l	100
18) Methyl Acetate	2.703	43	6686	1.159	ug/l	97
20) Methylene Chloride	2.788	84	13155	3.841	ug/l	94
37) Ethyl Acetate	4.715	43	73647	14.939	ug/l	99
43) Isopropyl Acetate	6.336	43	302465	36.216	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031623\
Data File : VX034543.D
Acq On : 16 Mar 2023 14:14
Operator : JC/MD
Sample : PB151426ZHE#05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB151426ZHE#05

Quant Time: Mar 17 06:25:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

