

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031623\
 Data File : VX034548.D
 Acq On : 16 Mar 2023 16:11
 Operator : JC/MD
 Sample : 01903-13DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE131D3-20230312DL

Quant Time: Mar 17 06:27:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	235302	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.763	114	405573	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	354518	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	146208	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	164385	48.021	ug/l	0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.040%
35) Dibromofluoromethane	5.385	113	127022	47.435	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.880%
50) Toluene-d8	8.647	98	487727	48.791	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.580%
62) 4-Bromofluorobenzene	11.079	95	184620	45.723	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.440%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	213895	81.778	ug/l	99
12) 1,1-Dichloroethene	2.325	96	1564	0.607	ug/l #	74
44) Trichloroethene	7.129	130	22865	7.314	ug/l	98
64) Tetrachloroethene	9.275	164	11437	4.803	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Mar 17 06:27:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

