

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031623\
 Data File : VX034550.D
 Acq On : 16 Mar 2023 16:58
 Operator : JC/MD
 Sample : 01903-18DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE131D2-20230312DL

Quant Time: Mar 17 06:28:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	238610	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	403940	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	353095	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	167422	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	163518	47.106	ug/l	0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.220%
35) Dibromofluoromethane	5.385	113	127351	47.751	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.500%
50) Toluene-d8	8.647	98	484559	48.670	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.340%
62) 4-Bromofluorobenzene	11.079	95	217577	54.103	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.200%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	131595	49.615	ug/l	99
12) 1,1-Dichloroethene	2.325	96	1413	0.541	ug/l #	93
27) cis-1,2-Dichloroethene	4.489	96	3361	1.006	ug/l	89
44) Trichloroethene	7.123	130	101667	32.653	ug/l	98
64) Tetrachloroethene	9.275	164	11447	4.826	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031623\
Data File : VX034550.D
Acq On : 16 Mar 2023 16:58
Operator : JC/MD
Sample : 01903-18DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE131D2-20230312DL

Quant Time: Mar 17 06:28:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

