

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031623\
 Data File : VX034552.D
 Acq On : 16 Mar 2023 17:45
 Operator : JC/MD
 Sample : 01904-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW8-MW01D3-20230312

Quant Time: Mar 17 06:28:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	232566	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	398541	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	402262	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	172130	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	160942	47.569	ug/l	0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.140%
35) Dibromofluoromethane	5.385	113	125272	47.607	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.220%
50) Toluene-d8	8.647	98	538244	54.794	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	109.580%
62) 4-Bromofluorobenzene	11.079	95	220651	55.611	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	111.220%
Target Compounds						
29) Tetrahydrofuran	5.001	42	414996	294.920	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031623\
Data File : VX034552.D
Acq On : 16 Mar 2023 17:45
Operator : JC/MD
Sample : 01904-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW8-MW01D3-20230312

Quant Time: Mar 17 06:28:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

