

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031623\
 Data File : VX034563.D
 Acq On : 17 Mar 2023 04:51
 Operator : JC/MD
 Sample : 01908-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 HILLBOR-DRUMS-0314

Quant Time: Mar 17 07:52:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	242730	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	413936	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	375989	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	181375	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	169551	48.015	ug/l	0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.020%
35) Dibromofluoromethane	5.385	113	130190	47.636	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.280%
50) Toluene-d8	8.647	98	512336	50.217	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.440%
62) 4-Bromofluorobenzene	11.079	95	236865	57.477	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	114.960%
Target Compounds						
					Qvalue	
16) Acetone	2.373	43	67728	40.717	ug/l	99
20) Methylene Chloride	2.794	84	11727	3.633	ug/l	95
37) Ethyl Acetate	4.714	43	55621	12.129	ug/l	98
43) Isopropyl Acetate	6.336	43	232731	29.957	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031623\
Data File : VX034563.D
Acq On : 17 Mar 2023 04:51
Operator : JC/MD
Sample : 01908-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
HILLBOR-DRUMS-0314

Quant Time: Mar 17 07:52:47 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

