

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031623\
 Data File : VX034564.D
 Acq On : 17 Mar 2023 05:14
 Operator : JC/MD
 Sample : 01905-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MAPLE-3

Quant Time: Mar 17 08:09:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	267535	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	460273	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	427178	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	209428	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	186105	47.816	ug/l	0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.640%
35) Dibromofluoromethane	5.385	113	145283	47.807	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.620%
50) Toluene-d8	8.647	98	569423	50.193	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.380%
62) 4-Bromofluorobenzene	11.079	95	268435	58.580	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	117.160%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.940	59	6203	7.653	ug/l #	87
16) Acetone	2.379	43	39549	21.572	ug/l	99
18) Methyl Acetate	2.702	43	6339	1.058	ug/l	95
20) Methylene Chloride	2.794	84	13311	3.741	ug/l	92
37) Ethyl Acetate	4.714	43	65661	12.877	ug/l	98
43) Isopropyl Acetate	6.342	43	274858	31.818	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031623\
Data File : VX034564.D
Acq On : 17 Mar 2023 05:14
Operator : JC/MD
Sample : 01905-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MAPLE-3

Quant Time: Mar 17 08:09:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

