

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031623\
 Data File : VX034566.D
 Acq On : 17 Mar 2023 06:01
 Operator : JC/MD
 Sample : 01907-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 33961

Quant Time: Mar 17 07:54:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	245158	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	422507	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	390212	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	174585	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	173487	48.643	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.280%
35) Dibromofluoromethane	5.385	113	134964	48.381	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.760%
50) Toluene-d8	8.647	98	523584	50.278	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.560%
62) 4-Bromofluorobenzene	11.079	95	211391	50.255	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.520%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.947	59	6501	8.753	ug/l #	88
16) Acetone	2.374	43	89794	53.448	ug/l	98
18) Methyl Acetate	2.703	43	5664	1.032	ug/l	100
20) Methylene Chloride	2.788	84	11594	3.556	ug/l	98
37) Ethyl Acetate	4.715	43	65675	14.031	ug/l	97
40) Benzene	6.038	78	20290	1.619	ug/l	97
43) Isopropyl Acetate	6.336	43	267441	33.727	ug/l	100
52) Toluene	8.720	92	12750	1.643	ug/l	96
69) o-Xylene	10.640	106	14598	2.471	ug/l	97
95) Naphthalene	13.780	128	24303	1.891	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031623\
Data File : VX034566.D
Acq On : 17 Mar 2023 06:01
Operator : JC/MD
Sample : 01907-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
33961

Quant Time: Mar 17 07:54:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

