

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031623\
 Data File : VX034567.D
 Acq On : 17 Mar 2023 06:25
 Operator : JC/MD
 Sample : 01907-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 33962

Quant Time: Mar 17 07:54:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	248749	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	423135	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	390893	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	174511	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	175417	48.474	ug/l	0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.940%
35) Dibromofluoromethane	5.385	113	134189	48.032	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.060%
50) Toluene-d8	8.647	98	524808	50.321	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.640%
62) 4-Bromofluorobenzene	11.079	95	208084	49.396	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.800%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.946	59	9313	12.358	ug/l #	83
16) Acetone	2.373	43	90875	53.311	ug/l	99
18) Methyl Acetate	2.703	43	5943	1.067	ug/l	97
20) Methylene Chloride	2.788	84	12438	3.760	ug/l	97
37) Ethyl Acetate	4.714	43	65606	13.995	ug/l	99
43) Isopropyl Acetate	6.336	43	266839	33.601	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031623\
Data File : VX034567.D
Acq On : 17 Mar 2023 06:25
Operator : JC/MD
Sample : 01907-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
33962

Quant Time: Mar 17 07:54:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

