

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031623\
 Data File : VX034559.D
 Acq On : 17 Mar 2023 02:30
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 17 07:49:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	81	0.00
2 T	Dichlorodifluoromethane	0.466	0.420	9.9	70	0.00
3 P	Chloromethane	0.839	0.626	25.4#	72	0.00
4 C	Vinyl Chloride	0.635	0.608	4.3#	76	0.00
5 T	Bromomethane	0.320	0.320	0.0	74	0.00
6 T	Chloroethane	0.365	0.395	-8.2	82	0.00
7 T	Trichlorofluoromethane	0.954	0.981	-2.8	79	0.00
8 T	Diethyl Ether	0.348	0.386	-10.9	88	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.556	0.520	6.5	72	0.00
10 T	Methyl Iodide	0.730	0.755	-3.4	82	0.00
11 T	Tert butyl alcohol	0.151	0.165	-9.3	99	0.00
12 CM	1,1-Dichloroethene	0.547	0.518	5.3#	76	0.00
13 T	Acrolein	0.120	0.142	-18.3	104	0.00
14 T	Allyl chloride	1.110	1.040	6.3	77	0.00
15 T	Acrylonitrile	0.327	0.359	-9.8	88	0.00
16 T	Acetone	0.343	0.318	7.3	75	0.00
17 T	Carbon Disulfide	1.540	1.218	20.9	65	0.00
18 T	Methyl Acetate	1.120	1.219	-8.8	90	0.00
19 T	Methyl tert-butyl Ether	1.960	2.059	-5.1	84	0.00
20 T	Methylene Chloride	0.665	0.637	4.2	81	0.00
21 T	trans-1,2-Dichloroethene	0.581	0.568	2.2	75	0.00
22 T	Diisopropyl ether	2.116	2.217	-4.8	82	0.00
23 T	Vinyl Acetate	1.610	1.721	-6.9	82	0.00
24 P	1,1-Dichloroethane	1.137	1.150	-1.1	80	0.00
25 T	2-Butanone	0.490	0.518	-5.7	84	0.00
26 T	2,2-Dichloropropane	0.989	0.790	20.1	63	0.00
27 T	cis-1,2-Dichloroethene	0.700	0.703	-0.4	80	0.00
28 T	Bromochloromethane	0.529	0.464	12.3	69	0.00
29 T	Tetrahydrofuran	0.303	0.332	-9.6	88	0.00
30 C	Chloroform	1.115	1.197	-7.4#	83	0.00
31 T	Cyclohexane	1.020	0.935	8.3	70	0.00
32 T	1,1,1-Trichloroethane	0.987	1.011	-2.4	79	0.00
33 S	1,2-Dichloroethane-d4	0.727	0.822	-13.1	103	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	84	0.00
35 S	Dibromofluoromethane	0.330	0.365	-10.6	100	0.00
36 T	1,1-Dichloropropene	0.507	0.477	5.9	74	0.00
37 T	Ethyl Acetate	0.554	0.584	-5.4	86	0.00
38 T	Carbon Tetrachloride	0.504	0.488	3.2	76	0.00
39 T	Methylcyclohexane	0.607	0.536	11.7	67	0.00
40 TM	Benzene	1.483	1.441	2.8	78	0.00
41 T	Methacrylonitrile	0.304	0.334	-9.9	90	0.00
42 TM	1,2-Dichloroethane	0.545	0.585	-7.3	86	0.00
43 T	Isopropyl Acetate	0.938	0.983	-4.8	84	0.00
44 TM	Trichloroethene	0.385	0.363	5.7	76	0.00
45 C	1,2-Dichloropropane	0.387	0.394	-1.8#	81	0.00
46 T	Dibromomethane	0.263	0.266	-1.1	82	0.00
47 T	Bromodichloromethane	0.510	0.524	-2.7	80	0.00
48 T	Methyl methacrylate	0.451	0.489	-8.4	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031623\
 Data File : VX034559.D
 Acq On : 17 Mar 2023 02:30
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 17 07:49:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.011	-22.2	96	0.00
50 S	Toluene-d8	1.232	1.322	-7.3	94	0.00
51 T	4-Methyl-2-Pentanone	0.544	0.612	-12.5	89	0.00
52 CM	Toluene	0.918	0.921	-0.3#	79	0.00
53 T	t-1,3-Dichloropropene	0.574	0.580	-1.0	77	0.00
54 T	cis-1,3-Dichloropropene	0.621	0.612	1.4	76	0.00
55 T	1,1,2-Trichloroethane	0.364	0.385	-5.8	82	0.00
56 T	Ethyl methacrylate	0.598	0.640	-7.0	83	0.00
57 T	1,3-Dichloropropane	0.648	0.681	-5.1	83	0.00
58 T	2-Chloroethyl Vinyl ether	0.308	0.315	-2.3	91	0.00
59 T	2-Hexanone	0.421	0.467	-10.9	88	0.00
60 T	Dibromochloromethane	0.382	0.405	-6.0	81	0.00
61 T	1,2-Dibromoethane	0.393	0.410	-4.3	82	0.00
62 S	4-Bromofluorobenzene	0.498	0.535	-7.4	96	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	86	0.00
64 T	Tetrachloroethene	0.336	0.323	3.9	78	0.00
65 PM	Chlorobenzene	1.097	1.052	4.1	80	0.00
66 T	1,1,1,2-Tetrachloroethane	0.400	0.402	-0.5	81	0.00
67 C	Ethyl Benzene	1.964	1.939	1.3#	80	0.00
68 T	m/p-Xylenes	0.754	0.727	3.6	78	0.00
69 T	o-Xylene	0.757	0.734	3.0	80	0.00
70 T	Styrene	1.222	1.221	0.1	81	0.00
71 P	Bromoform	0.308	0.307	0.3	79	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	85	0.00
73 T	Isopropylbenzene	3.864	3.807	1.5	79	0.00
74 T	N-amyl acetate	1.851	1.941	-4.9	85	0.00
75 P	1,1,2,2-Tetrachloroethane	1.290	1.334	-3.4	84	0.00
76 T	1,2,3-Trichloropropane	1.148	1.121	2.4	84	0.00
77 T	Bromobenzene	0.927	0.914	1.4	80	0.00
78 T	n-propylbenzene	4.463	4.449	0.3	77	0.00
79 T	2-Chlorotoluene	2.756	2.747	0.3	80	0.00
80 T	1,3,5-Trimethylbenzene	3.241	3.288	-1.5	80	0.00
81 T	trans-1,4-Dichloro-2-butene	0.424	0.376	11.3	71	0.00
82 T	4-Chlorotoluene	3.175	3.218	-1.4	80	0.00
83 T	tert-Butylbenzene	3.268	3.261	0.2	80	0.00
84 T	1,2,4-Trimethylbenzene	3.313	3.310	0.1	81	0.00
85 T	sec-Butylbenzene	3.995	4.004	-0.2	78	0.00
86 T	p-Isopropyltoluene	3.358	3.359	-0.0	78	0.00
87 T	1,3-Dichlorobenzene	1.817	1.759	3.2	79	0.00
88 T	1,4-Dichlorobenzene	1.837	1.780	3.1	80	0.00
89 T	n-Butylbenzene	2.985	2.934	1.7	74	0.00
90 T	Hexachloroethane	0.621	0.609	1.9	75	0.00
91 T	1,2-Dichlorobenzene	1.756	1.749	0.4	81	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.293	0.304	-3.8	82	0.00
93 T	1,2,4-Trichlorobenzene	1.098	1.056	3.8	73	0.00
94 T	Hexachlorobutadiene	0.442	0.411	7.0	71	0.00
95 T	Naphthalene	3.681	3.744	-1.7	80	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031623\
Data File : VX034559.D
Acq On : 17 Mar 2023 02:30
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Mar 17 07:49:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.073	1.052	2.0	77	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6