

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031623\
 Data File : VX034534.D
 Acq On : 16 Mar 2023 10:40
 Operator : JC/MD
 Sample : VX0316WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0316WBL01

Quant Time: Mar 17 06:20:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	256998	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	440204	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	389716	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	175901	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	176940	47.325	ug/l	0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.660%
35) Dibromofluoromethane	5.385	113	138339	47.597	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.200%
50) Toluene-d8	8.647	98	529598	48.811	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.620%
62) 4-Bromofluorobenzene	11.079	95	210937	48.131	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.260%


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031623\
Data File : VX034534.D
Acq On : 16 Mar 2023 10:40
Operator : JC/MD
Sample : VX0316WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0316WBL01

Quant Time: Mar 17 06:20:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 711
Delta R.T. 0.006 min
Lab File: VX034534.D
Acq: 16 Mar 2023 10:40




Instrument :
MSVOA_X
ClientSampleId :
VX0316WBL01


Tgt Ion:168 Resp: 256998
Ion Ratio Lower Upper
168 100
99 59.9 45.7 68.5



#33
1,2-Dichloroethane-d4
Concen: 47.325 ug/l
RT: 5.958 min Scan# 799
Delta R.T. 0.012 min
Lab File: VX034534.D
Acq: 16 Mar 2023 10:40

Tgt Ion: 65 Resp: 176940
Ion Ratio Lower Upper
65 100
67 51.7 0.0 104.0









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 91
Delta R.T. 0.000 min
Lab File: VX034534.D
Acq: 16 Mar 2023 10:40




Instrument :
MSVOA_X
ClientSampleId :
VX0316WBL01


Tgt Ion:114 Resp: 440204
Ion Ratio Lower Upper
114 100
63 21.2 0.0 42.2
88 17.1 0.0 33.2



#35
Dibromofluoromethane
Concen: 47.597 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.006 min
Lab File: VX034534.D
Acq: 16 Mar 2023 10:40

Tgt Ion:113 Resp: 138339
Ion Ratio Lower Upper
113 100
111 105.2 83.7 125.5
192 19.7 15.4 23.2









#50
Toluene-d8
Concen: 48.811 ug/l
RT: 8.647 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX034534.D
Acq: 16 Mar 2023 10:40




Instrument :
MSVOA_X
ClientSampleId :
VX0316WBL01


Tgt Ion: 98 Resp: 529598
Ion Ratio Lower Upper
98 100
100 65.4 52.5 78.7



#62
4-Bromofluorobenzene
Concen: 48.131 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX034534.D
Acq: 16 Mar 2023 10:40

Tgt Ion: 95 Resp: 210937
Ion Ratio Lower Upper
95 100
174 74.9 0.0 147.4
176 72.1 0.0 140.6





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.055 min Scan# 1471

Delta R.T. 0.000 min

Lab File: VX034534.D

Acq: 16 Mar 2023 10:40

Instrument :

MSVOA_X

ClientSampleId :

VX0316WBL01





Tgt Ion:117 Resp: 389716

Ion Ratio Lower Upper

117 100

82 55.7 45.0 67.4

119 33.0 26.0 39.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.024 min Scan# 1794

Delta R.T. 0.000 min

Lab File: VX034534.D

Acq: 16 Mar 2023 10:40

Tgt Ion:152 Resp: 175901

Ion Ratio Lower Upper

152 100

115 61.5 45.0 135.0

150 158.8 0.0 352.4

