

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031720\
 Data File : VX015314.D
 Acq On : 17 Mar 2020 11:07
 Operator : JC/SP
 Sample : L1919-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Mar 18 02:07:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 04 14:19:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	299367	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	485357	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	469187	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	232573	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	174722	49.20	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.40%	
35) Dibromofluoromethane	5.49	113	149209	48.73	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.46%	
50) Toluene-d8	8.71	98	587242	50.42	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.84%	
62) 4-Bromofluorobenzene	11.13	95	215541	51.88	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.76%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031720\
Data File : VX015314.D
Acq On : 17 Mar 2020 11:07
Operator : JC/SP
Sample : L1919-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Mar 18 02:07:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 04 14:19:10 2020
Response via : Initial Calibration

