

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031720\
 Data File : VX015316.D
 Acq On : 17 Mar 2020 11:52
 Operator : JC/SP
 Sample : L1919-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANA

Quant Time: Mar 18 02:13:15 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 04 14:19:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	303132	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	491250	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	475504	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	237085	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	177581	49.38	ug/l	0.00
Spiked Amount						
			Recovery	=	98.76%	
35) Dibromofluoromethane	5.49	113	151596	48.92	ug/l	0.00
Spiked Amount						
			Recovery	=	97.84%	
50) Toluene-d8	8.71	98	600182	50.91	ug/l	0.00
Spiked Amount						
			Recovery	=	101.82%	
62) 4-Bromofluorobenzene	11.13	95	219672	52.24	ug/l	0.00
Spiked Amount						
			Recovery	=	104.48%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031720\
Data File : VX015316.D
Acq On : 17 Mar 2020 11:52
Operator : JC/SP
Sample : L1919-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANA

Quant Time: Mar 18 02:13:15 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 04 14:19:10 2020
Response via : Initial Calibration

