


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX031720\
 Data File : VX015308.D
 Acq On : 17 Mar 2020 08:12
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
 Title : SW846 8260
 Last Update : Wed Mar 04 14:19:10 2020



AutoFind: Scans 1679, 1680, 1681; Background Corrected with Scan 1672

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.3	15152	PASS
75	95	30	60	50.3	41741	PASS
95	95	100	100	100.0	82987	PASS
96	95	5	9	6.4	5333	PASS
173	174	0.00	2	1.6	1094	PASS
174	95	50	100	84.7	70269	PASS
175	174	5	9	7.7	5440	PASS
176	174	95	101	97.2	68281	PASS
177	176	5	9	7.1	4877	PASS