

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031721\
 Data File : VX021267.D
 Acq On : 17 Mar 2021 14:00
 Operator : JC/MD
 Sample : M1663-03 5000X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1875

Quant Time: Mar 18 03:53:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 16 14:00:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.617	168	112219	50.00	ug/l	-0.02
34) 1,4-Difluorobenzene	6.823	114	185774	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	168229	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	83813	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	88357	48.00	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.00%
35) Dibromofluoromethane	5.453	113	62303	49.29	ug/l	-0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.58%
50) Toluene-d8	8.694	98	230199	50.19	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.38%
62) 4-Bromofluorobenzene	11.118	95	91635	51.44	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.88%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031721\
Data File : VX021267.D
Acq On : 17 Mar 2021 14:00
Operator : JC/MD
Sample : M1663-03 5000X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1875

Quant Time: Mar 18 03:53:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 16 14:00:21 2021
Response via : Initial Calibration

