

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031721\
 Data File : VX021260.D
 Acq On : 17 Mar 2021 09:51
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 18 03:49:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 16 14:00:21 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	107	-0.01
2 T	Dichlorodifluoromethane	50.000	49.684	0.6	103	0.00
3 P	Chloromethane	50.000	45.032	9.9	94	0.00
4 C	Vinyl Chloride	50.000	47.187	5.6#	97	0.00
5 T	Bromomethane	50.000	49.404	1.2	108	0.00
6 T	Chloroethane	50.000	47.584	4.8	98	0.00
7 T	Trichlorofluoromethane	50.000	50.004	-0.0	99	0.00
8 T	Diethyl Ether	50.000	47.919	4.2	98	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	49.292	1.4	102	0.00
10 T	Methyl Iodide	50.000	45.380	9.2	97	0.00
11 T	Tert butyl alcohol	250.000	226.413	9.4	97	0.00
12 CM	1,1-Dichloroethene	50.000	48.097	3.8#	101	0.00
13 T	Acrolein	250.000	238.489	4.6	104	0.00
14 T	Allyl chloride	50.000	48.223	3.6	99	0.00
15 T	Acrylonitrile	250.000	240.083	4.0	99	0.00
16 T	Acetone	250.000	214.240	14.3	92	0.00
17 T	Carbon Disulfide	50.000	46.280	7.4	100	0.00
18 T	Methyl Acetate	50.000	45.082	9.8	98	0.00
19 T	Methyl tert-butyl Ether	50.000	49.959	0.1	101	0.00
20 T	Methylene Chloride	50.000	46.892	6.2	101	0.00
21 T	trans-1,2-Dichloroethene	50.000	47.728	4.5	100	0.00
22 T	Diisopropyl ether	50.000	49.637	0.7	100	0.00
23 T	Vinyl Acetate	250.000	253.257	-1.3	99	0.00
24 P	1,1-Dichloroethane	50.000	48.640	2.7	100	0.00
25 T	2-Butanone	250.000	236.626	5.3	97	0.00
26 T	2,2-Dichloropropane	50.000	51.671	-3.3	104	0.00
27 T	cis-1,2-Dichloroethene	50.000	48.587	2.8	101	0.00
28 T	Bromochloromethane	50.000	48.593	2.8	102	0.00
29 T	Tetrahydrofuran	250.000	233.364	6.7	96	0.00
30 C	Chloroform	50.000	49.526	0.9#	100	0.00
31 T	Cyclohexane	50.000	48.381	3.2	100	0.00
32 T	1,1,1-Trichloroethane	50.000	49.637	0.7	100	0.00
33 S	1,2-Dichloroethane-d4	50.000	47.652	4.7	104	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	102	0.00
35 S	Dibromofluoromethane	50.000	49.808	0.4	104	0.00
36 T	1,1-Dichloropropene	50.000	50.915	-1.8	103	0.00
37 T	Ethyl Acetate	50.000	46.363	7.3	96	0.00
38 T	Carbon Tetrachloride	50.000	52.794	-5.6	102	0.00
39 T	Methylcyclohexane	50.000	53.109	-6.2	104	0.00
40 TM	Benzene	50.000	51.837	-3.7	101	0.00
41 T	Methacrylonitrile	50.000	46.762	6.5	99	0.00
42 TM	1,2-Dichloroethane	50.000	50.900	-1.8	101	0.00
43 T	Isopropyl Acetate	50.000	49.243	1.5	98	0.00
44 TM	Trichloroethene	50.000	51.354	-2.7	102	0.00
45 C	1,2-Dichloropropane	50.000	52.333	-4.7#	102	0.00
46 T	Dibromomethane	50.000	50.643	-1.3	101	0.00
47 T	Bromodichloromethane	50.000	52.710	-5.4	102	0.00
48 T	Methyl methacrylate	50.000	48.448	3.1	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031721\
 Data File : VX021260.D
 Acq On : 17 Mar 2021 09:51
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 18 03:49:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 16 14:00:21 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	924.992	7.5	94	0.00
50 S	Toluene-d8	50.000	49.543	0.9	104	0.00
51 T	4-Methyl-2-Pentanone	250.000	251.771	-0.7	98	0.00
52 CM	Toluene	50.000	52.161	-4.3#	100	0.00
53 T	t-1,3-Dichloropropene	50.000	52.531	-5.1	102	0.00
54 T	cis-1,3-Dichloropropene	50.000	52.532	-5.1	101	0.00
55 T	1,1,2-Trichloroethane	50.000	52.483	-5.0	101	0.00
56 T	Ethyl methacrylate	50.000	52.324	-4.6	101	0.00
57 T	1,3-Dichloropropane	50.000	52.132	-4.3	102	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	242.646	2.9	98	0.00
59 T	2-Hexanone	250.000	252.020	-0.8	97	0.00
60 T	Dibromochloromethane	50.000	52.876	-5.8	100	0.00
61 T	1,2-Dibromoethane	50.000	51.718	-3.4	102	0.00
62 S	4-Bromofluorobenzene	50.000	51.032	-2.1	106	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	100	0.00
64 T	Tetrachloroethene	50.000	50.910	-1.8	104	0.00
65 PM	Chlorobenzene	50.000	50.992	-2.0	103	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	51.906	-3.8	103	0.00
67 C	Ethyl Benzene	50.000	51.799	-3.6#	102	0.00
68 T	m/p-Xylenes	100.000	106.328	-6.3	104	0.00
69 T	o-Xylene	50.000	52.422	-4.8	103	0.00
70 T	Styrene	50.000	53.538	-7.1	105	0.00
71 P	Bromoform	50.000	51.629	-3.3	99	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	102	0.00
73 T	Isopropylbenzene	50.000	52.668	-5.3	104	0.00
74 T	N-amyl acetate	50.000	50.278	-0.6	99	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	50.594	-1.2	99	0.00
76 T	1,2,3-Trichloropropane	50.000	53.410	-6.8	115	0.00
77 T	Bromobenzene	50.000	52.672	-5.3	104	0.00
78 T	n-propylbenzene	50.000	53.434	-6.9	105	0.00
79 T	2-Chlorotoluene	50.000	51.788	-3.6	103	0.00
80 T	1,3,5-Trimethylbenzene	50.000	52.320	-4.6	101	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	50.517	-1.0	101	0.00
82 T	4-Chlorotoluene	50.000	51.475	-3.0	102	0.00
83 T	tert-Butylbenzene	50.000	53.309	-6.6	105	0.00
84 T	1,2,4-Trimethylbenzene	50.000	52.487	-5.0	102	0.00
85 T	sec-Butylbenzene	50.000	53.167	-6.3	104	0.00
86 T	p-Isopropyltoluene	50.000	53.324	-6.6	103	0.00
87 T	1,3-Dichlorobenzene	50.000	51.931	-3.9	105	0.00
88 T	1,4-Dichlorobenzene	50.000	50.772	-1.5	102	0.00
89 T	n-Butylbenzene	50.000	54.071	-8.1	106	0.00
90 T	Hexachloroethane	50.000	55.730	-11.5	107	0.00
91 T	1,2-Dichlorobenzene	50.000	50.873	-1.7	101	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	48.157	3.7	97	0.00
93 T	1,2,4-Trichlorobenzene	50.000	51.064	-2.1	102	0.00
94 T	Hexachlorobutadiene	50.000	53.466	-6.9	104	0.00
95 T	Naphthalene	50.000	49.666	0.7	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031721\
Data File : VX021260.D
Acq On : 17 Mar 2021 09:51
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Mar 18 03:49:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 16 14:00:21 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	52.191	-4.4	100	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6