

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031722\
 Data File : VX027492.D
 Acq On : 17 Mar 2022 17:17
 Operator : JC/MD
 Sample : VX0317WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0317WBL01

Quant Time: Mar 18 02:05:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	347936	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	570099	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	507537	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	232663	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	215329	50.237	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	100.480%
35) Dibromofluoromethane	5.391	113	189327	51.617	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.240%
50) Toluene-d8	8.653	98	660755	47.606	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.220%
62) 4-Bromofluorobenzene	11.079	95	238944	46.837	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	93.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031722\
Data File : VX027492.D
Acq On : 17 Mar 2022 17:17
Operator : JC/MD
Sample : VX0317WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0317WBL01

Quant Time: Mar 18 02:05:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

