

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031722\
 Data File : VX027497.D
 Acq On : 17 Mar 2022 19:22
 Operator : JC/MD
 Sample : N1997-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-103

Quant Time: Mar 18 02:06:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	334854	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	545834	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	495673	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	226503	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	208549	50.557	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 101.120%	
35) Dibromofluoromethane	5.391	113	183501	52.253	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 104.500%	
50) Toluene-d8	8.653	98	641513	48.274	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 96.540%	
62) 4-Bromofluorobenzene	11.085	95	231907	47.479	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 94.960%	
Target Compounds						Qvalue
16) Acetone	2.386	43	5568	3.312	ug/l	90
19) Methyl tert-butyl Ether	3.117	73	43302	4.128	ug/l	98
30) Chloroform	5.105	83	3698	0.580	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031722\
Data File : VX027497.D
Acq On : 17 Mar 2022 19:22
Operator : JC/MD
Sample : N1997-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-103

Quant Time: Mar 18 02:06:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

