

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031722\
 Data File : VX027500.D
 Acq On : 17 Mar 2022 20:33
 Operator : JC/MD
 Sample : N1997-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-105

Quant Time: Mar 18 02:07:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	340204	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	556511	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	503645	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	228274	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	210374	50.197	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	=	100.400%	
35) Dibromofluoromethane	5.391	113	183020	51.116	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	=	102.240%	
50) Toluene-d8	8.653	98	648050	47.830	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	=	95.660%	
62) 4-Bromofluorobenzene	11.085	95	235829	47.355	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	=	94.720%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	5043	2.952	ug/l	99
19) Methyl tert-butyl Ether	3.117	73	6318	0.593	ug/l	94
20) Methylene Chloride	2.788	84	1180	0.321	ug/l #	90
30) Chloroform	5.099	83	13498	2.084	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031722\
Data File : VX027500.D
Acq On : 17 Mar 2022 20:33
Operator : JC/MD
Sample : N1997-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-105

Quant Time: Mar 18 02:07:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

