

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031722\
 Data File : VX027505.D
 Acq On : 17 Mar 2022 22:31
 Operator : JC/MD
 Sample : N1997-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-111

Quant Time: Mar 18 02:08:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	352029	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	541700	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	466522	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	209718	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	204288	47.107	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.220%
35) Dibromofluoromethane	5.391	113	187945	53.927	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	107.860%
50) Toluene-d8	8.653	98	614520	46.596	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	93.200%
62) 4-Bromofluorobenzene	11.085	95	218478	45.071	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.140%
Target Compounds						
16) Acetone	2.386	43	5781	3.270	ug/l	99
30) Chloroform	5.105	83	2542	0.379	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031722\
Data File : VX027505.D
Acq On : 17 Mar 2022 22:31
Operator : JC/MD
Sample : N1997-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-111

Quant Time: Mar 18 02:08:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

