

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031722\
 Data File : VX027512.D
 Acq On : 18 Mar 2022 01:16
 Operator : JC/MD
 Sample : N1997-18
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-122

Quant Time: Mar 18 03:41:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	333511	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	552134	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	486340	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	209609	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	208031	50.634	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 101.260%	
35) Dibromofluoromethane	5.391	113	184327	51.889	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 103.780%	
50) Toluene-d8	8.653	98	642486	47.795	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 95.600%	
62) 4-Bromofluorobenzene	11.085	95	220814	44.692	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 89.380%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	6696	3.998	ug/l	100
20) Methylene Chloride	2.788	84	999	0.277	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031722\
Data File : VX027512.D
Acq On : 18 Mar 2022 01:16
Operator : JC/MD
Sample : N1997-18
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-122

Quant Time: Mar 18 03:41:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

