

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031722\
 Data File : VX027515.D
 Acq On : 18 Mar 2022 02:51
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 18 03:42:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	96	0.00
2 T	Dichlorodifluoromethane	0.480	0.537	-11.9	97	0.00
3 P	Chloromethane	0.426	0.451	-5.9	95	0.00
4 C	Vinyl Chloride	0.492	0.557	-13.2#	101	0.00
5 T	Bromomethane	0.272	0.197	27.6#	73	0.00
6 T	Chloroethane	0.302	0.327	-8.3	96	0.00
7 T	Trichlorofluoromethane	0.813	0.860	-5.8	96	0.00
8 T	Diethyl Ether	0.303	0.320	-5.6	95	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.484	0.539	-11.4	101	0.00
10 T	Methyl Iodide	0.641	0.533	16.8	73	0.00
11 T	Tert butyl alcohol	0.117	0.128	-9.4	94	0.00
12 CM	1,1-Dichloroethene	0.464	0.533	-14.9#	106	0.00
13 T	Acrolein	0.091	0.075	17.6	75	0.00
14 T	Allyl chloride	0.727	0.797	-9.6	99	0.00
15 T	Acrylonitrile	0.267	0.312	-16.9	104	0.00
16 T	Acetone	0.251	0.266	-6.0	100	0.00
17 T	Carbon Disulfide	1.181	1.338	-13.3	104	0.00
18 T	Methyl Acetate	0.602	0.689	-14.5	106	0.00
19 T	Methyl tert-butyl Ether	1.566	1.759	-12.3	102	0.00
20 T	Methylene Chloride	0.540	0.593	-9.8	106	0.00
21 T	trans-1,2-Dichloroethene	0.495	0.557	-12.5	104	0.00
22 T	Diisopropyl ether	1.419	1.619	-14.1	103	0.00
23 T	Vinyl Acetate	1.234	1.419	-15.0	102	0.00
24 P	1,1-Dichloroethane	0.863	0.971	-12.5	102	0.00
25 T	2-Butanone	0.378	0.430	-13.8	102	0.00
26 T	2,2-Dichloropropane	0.705	0.597	15.3	76	0.00
27 T	cis-1,2-Dichloroethene	0.574	0.651	-13.4	103	0.00
28 T	Bromochloromethane	0.348	0.334	4.0	85	0.00
29 T	Tetrahydrofuran	0.242	0.283	-16.9	104	0.00
30 C	Chloroform	0.952	1.064	-11.8#	102	0.00
31 T	Cyclohexane	0.765	0.855	-11.8	101	0.00
32 T	1,1,1-Trichloroethane	0.867	0.956	-10.3	101	0.00
33 S	1,2-Dichloroethane-d4	0.616	0.578	6.2	86	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
35 S	Dibromofluoromethane	0.322	0.309	4.0	87	0.00
36 T	1,1-Dichloropropene	0.440	0.479	-8.9	102	0.00
37 T	Ethyl Acetate	0.438	0.516	-17.8	104	0.00
38 T	Carbon Tetrachloride	0.503	0.542	-7.8	100	0.00
39 T	Methylcyclohexane	0.553	0.584	-5.6	99	0.00
40 TM	Benzene	1.261	1.409	-11.7	104	0.00
41 T	Methacrylonitrile	0.239	0.278	-16.3	102	0.00
42 TM	1,2-Dichloroethane	0.473	0.523	-10.6	100	0.00
43 T	Isopropyl Acetate	0.705	0.786	-11.5	102	0.00
44 TM	Trichloroethene	0.383	0.406	-6.0	101	0.00
45 C	1,2-Dichloropropane	0.321	0.355	-10.6#	103	0.00
46 T	Dibromomethane	0.245	0.266	-8.6	101	0.00
47 T	Bromodichloromethane	0.462	0.509	-10.2	100	0.00
48 T	Methyl methacrylate	0.349	0.399	-14.3	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031722\
 Data File : VX027515.D
 Acq On : 18 Mar 2022 02:51
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 18 03:42:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.010	-11.1	94	0.00
50 S	Toluene-d8	1.217	1.131	7.1	82	0.00
51 T	4-Methyl-2-Pentanone	0.450	0.519	-15.3	99	0.00
52 CM	Toluene	0.845	0.927	-9.7#	96	0.00
53 T	t-1,3-Dichloropropene	0.474	0.508	-7.2	91	0.00
54 T	cis-1,3-Dichloropropene	0.516	0.562	-8.9	97	0.00
55 T	1,1,2-Trichloroethane	0.344	0.381	-10.8	94	0.00
56 T	Ethyl methacrylate	0.515	0.574	-11.5	93	0.00
57 T	1,3-Dichloropropane	0.559	0.617	-10.4	96	0.00
58 T	2-Chloroethyl Vinyl ether	0.242	0.275	-13.6	97	0.00
59 T	2-Hexanone	0.347	0.399	-15.0	99	0.00
60 T	Dibromochloromethane	0.379	0.417	-10.0	94	0.00
61 T	1,2-Dibromoethane	0.378	0.411	-8.7	94	0.00
62 S	4-Bromofluorobenzene	0.447	0.419	6.3	84	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	90	0.00
64 T	Tetrachloroethene	0.389	0.414	-6.4	94	0.00
65 PM	Chlorobenzene	1.029	1.120	-8.8	94	0.00
66 T	1,1,1,2-Tetrachloroethane	0.384	0.426	-10.9	94	0.00
67 C	Ethyl Benzene	1.761	1.977	-12.3#	97	0.00
68 T	m/p-Xylenes	0.693	0.769	-11.0	97	0.00
69 T	o-Xylene	0.673	0.775	-15.2	99	0.00
70 T	Styrene	1.099	1.280	-16.5	98	0.00
71 P	Bromoform	0.307	0.349	-13.7	96	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
73 T	Isopropylbenzene	3.642	3.863	-6.1	97	0.00
74 T	N-amyl acetate	1.291	1.384	-7.2	95	0.00
75 P	1,1,2,2-Tetrachloroethane	1.154	1.240	-7.5	101	0.00
76 T	1,2,3-Trichloropropane	1.046	1.124	-7.5	96	0.00
77 T	Bromobenzene	0.918	0.964	-5.0	96	0.00
78 T	n-propylbenzene	3.998	4.298	-7.5	97	0.00
79 T	2-Chlorotoluene	2.508	2.614	-4.2	96	0.00
80 T	1,3,5-Trimethylbenzene	3.065	3.262	-6.4	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.344	0.339	1.5	87	0.00
82 T	4-Chlorotoluene	2.855	2.975	-4.2	96	0.00
83 T	tert-Butylbenzene	3.036	3.290	-8.4	98	0.00
84 T	1,2,4-Trimethylbenzene	3.057	3.349	-9.6	99	0.00
85 T	sec-Butylbenzene	3.669	4.029	-9.8	98	0.00
86 T	p-Isopropyltoluene	3.152	3.461	-9.8	98	0.00
87 T	1,3-Dichlorobenzene	1.701	1.822	-7.1	99	0.00
88 T	1,4-Dichlorobenzene	1.721	1.835	-6.6	100	0.00
89 T	n-Butylbenzene	2.630	2.749	-4.5	93	0.00
90 T	Hexachloroethane	0.503	0.524	-4.2	92	0.00
91 T	1,2-Dichlorobenzene	1.661	1.775	-6.9	99	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.265	0.293	-10.6	102	0.00
93 T	1,2,4-Trichlorobenzene	1.064	1.101	-3.5	97	0.00
94 T	Hexachlorobutadiene	0.470	0.472	-0.4	93	0.00
95 T	Naphthalene	3.465	3.829	-10.5	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031722\
Data File : VX027515.D
Acq On : 18 Mar 2022 02:51
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Mar 18 03:42:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.077	1.118	-3.8	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6