

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031723\
 Data File : VX034580.D
 Acq On : 17 Mar 2023 13:35
 Operator : JC/MD
 Sample : PB151466ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB151466ZHE#01

Quant Time: Mar 20 00:00:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	237216	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	407059	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	417438	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	179387	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	165405	47.929	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.860%
35) Dibromofluoromethane	5.385	113	127902	47.590	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.180%
50) Toluene-d8	8.647	98	567338	56.547	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	113.100%#
62) 4-Bromofluorobenzene	11.079	95	215640	53.211	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.420%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	81190	49.945	ug/l	95
20) Methylene Chloride	2.788	84	11442	3.627	ug/l	96
37) Ethyl Acetate	4.709	43	69698	15.455	ug/l	98
43) Isopropyl Acetate	6.342	43	286109	37.450	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031723\
Data File : VX034580.D
Acq On : 17 Mar 2023 13:35
Operator : JC/MD
Sample : PB151466ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB151466ZHE#01

Quant Time: Mar 20 00:00:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

