

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031723\
 Data File : VX034586.D
 Acq On : 17 Mar 2023 15:57
 Operator : JC/MD
 Sample : PB151466ZHE#07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB151466ZHE#07

Quant Time: Mar 20 00:03:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	232469	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.763	114	398099	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	356186	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	164027	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	161645	47.796	ug/l	0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.600%
35) Dibromofluoromethane	5.385	113	124599	47.404	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.800%
50) Toluene-d8	8.647	98	488853	49.821	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.640%
62) 4-Bromofluorobenzene	11.079	95	207585	52.376	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.760%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	40293	25.293	ug/l	98
20) Methylene Chloride	2.788	84	15139	4.896	ug/l	96
37) Ethyl Acetate	4.714	43	66055	14.977	ug/l	98
43) Isopropyl Acetate	6.342	43	268018	35.872	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031723\
Data File : VX034586.D
Acq On : 17 Mar 2023 15:57
Operator : JC/MD
Sample : PB151466ZHE#07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB151466ZHE#07

Quant Time: Mar 20 00:03:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

