

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031723\
 Data File : VX034587.D
 Acq On : 17 Mar 2023 16:20
 Operator : JC/MD
 Sample : PB151466ZHE#08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB151466ZHE#08

Quant Time: Mar 20 00:03:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	233654	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	397660	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	359373	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	167725	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	163400	48.070	ug/l	0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.140%
35) Dibromofluoromethane	5.391	113	124895	47.569	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.140%
50) Toluene-d8	8.647	98	484532	49.435	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.880%
62) 4-Bromofluorobenzene	11.079	95	167669	42.352	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	84.700%
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	40335	25.191	ug/l	97
20) Methylene Chloride	2.788	84	15317	4.929	ug/l	97
37) Ethyl Acetate	4.714	43	68160	15.471	ug/l	99
43) Isopropyl Acetate	6.342	43	270259	36.212	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031723\
Data File : VX034587.D
Acq On : 17 Mar 2023 16:20
Operator : JC/MD
Sample : PB151466ZHE#08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB151466ZHE#08

Quant Time: Mar 20 00:03:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

