

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031723\
 Data File : VX034588.D
 Acq On : 17 Mar 2023 16:44
 Operator : JC/MD
 Sample : PB151466ZHE#09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB151466ZHE#09

Quant Time: Mar 20 00:04:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	220182	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.763	114	378100	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	337698	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	137529	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	152022	47.459	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.920%
35) Dibromofluoromethane	5.385	113	118802	47.589	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.180%
50) Toluene-d8	8.647	98	465323	49.932	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.860%
62) 4-Bromofluorobenzene	11.079	95	159521	42.378	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	84.760%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	39648	26.277	ug/l	97
20) Methylene Chloride	2.788	84	14978	5.115	ug/l	94
25) 2-Butanone	4.562	43	11351	5.258	ug/l #	88
37) Ethyl Acetate	4.715	43	67419	16.095	ug/l	100
43) Isopropyl Acetate	6.342	43	269301	37.950	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031723\
Data File : VX034588.D
Acq On : 17 Mar 2023 16:44
Operator : JC/MD
Sample : PB151466ZHE#09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB151466ZHE#09

Quant Time: Mar 20 00:04:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

