

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031723\
 Data File : VX034594.D
 Acq On : 17 Mar 2023 19:06
 Operator : JC/MD
 Sample : PB151466ZHE#15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB151466ZHE#15

Quant Time: Mar 20 00:07:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	218965	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.763	114	372011	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	337205	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	138899	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	151301	47.497	ug/l	0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.000%
35) Dibromofluoromethane	5.385	113	117258	47.740	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.480%
50) Toluene-d8	8.647	98	456624	49.800	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.600%
62) 4-Bromofluorobenzene	11.079	95	161292	43.550	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.100%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	34205	22.795	ug/l	96
18) Methyl Acetate	2.703	43	5759	1.175	ug/l	96
20) Methylene Chloride	2.788	84	11762	4.039	ug/l	92
25) 2-Butanone	4.574	43	11079	5.160	ug/l	92
37) Ethyl Acetate	4.715	43	67365	16.345	ug/l	98
43) Isopropyl Acetate	6.342	43	268843	38.505	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031723\
Data File : VX034594.D
Acq On : 17 Mar 2023 19:06
Operator : JC/MD
Sample : PB151466ZHE#15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB151466ZHE#15

Quant Time: Mar 20 00:07:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

