

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031723\
 Data File : VX034596.D
 Acq On : 17 Mar 2023 19:53
 Operator : JC/MD
 Sample : 01928-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 031423-1

Quant Time: Mar 20 00:08:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	220567	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	379381	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	354746	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	155650	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	155380	48.423	ug/l	0.01
Spiked Amount	50.000	Range 78 - 117	Recovery	=	96.840%	
35) Dibromofluoromethane	5.391	113	120578	48.138	ug/l	0.01
Spiked Amount	50.000	Range 75 - 124	Recovery	=	96.280%	
50) Toluene-d8	8.647	98	471558	50.430	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	100.860%	
62) 4-Bromofluorobenzene	11.079	95	187401	49.616	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	99.240%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	94182	62.310	ug/l	100
18) Methyl Acetate	2.703	43	5598	1.133	ug/l	98
20) Methylene Chloride	2.788	84	19256	6.564	ug/l	97
25) 2-Butanone	4.574	43	10956	5.066	ug/l #	83
37) Ethyl Acetate	4.715	43	62593	14.892	ug/l	99
43) Isopropyl Acetate	6.342	43	259180	36.400	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031723\
Data File : VX034596.D
Acq On : 17 Mar 2023 19:53
Operator : JC/MD
Sample : 01928-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
031423-1

Quant Time: Mar 20 00:08:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

