

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031723\
 Data File : VX034597.D
 Acq On : 17 Mar 2023 20:17
 Operator : JC/MD
 Sample : 01932-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CE-86B-031423

Quant Time: Mar 20 00:08:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	225888	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	389575	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	349541	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	158967	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	157028	47.784	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.560%
35) Dibromofluoromethane	5.385	113	122725	47.713	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.420%
50) Toluene-d8	8.647	98	475890	49.561	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.120%
62) 4-Bromofluorobenzene	11.079	95	179280	46.224	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.440%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	115115	74.365	ug/l	99
20) Methylene Chloride	2.788	84	19267	6.413	ug/l	92
37) Ethyl Acetate	4.721	43	62739	14.536	ug/l	98
43) Isopropyl Acetate	6.342	43	252551	34.541	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031723\
Data File : VX034597.D
Acq On : 17 Mar 2023 20:17
Operator : JC/MD
Sample : 01932-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CE-86B-031423

Quant Time: Mar 20 00:08:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

