

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031723\
 Data File : VX034605.D
 Acq On : 18 Mar 2023 02:32
 Operator : JC/MD
 Sample : 01932-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CE-43-031423

Quant Time: Mar 20 02:07:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	221296	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.763	114	378010	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	342375	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	144263	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	156510	48.614	ug/l	0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.220%
35) Dibromofluoromethane	5.391	113	121602	48.723	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.440%
50) Toluene-d8	8.647	98	465437	49.956	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.920%
62) 4-Bromofluorobenzene	11.079	95	173706	46.157	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.320%
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	81878	53.991	ug/l	97
20) Methylene Chloride	2.788	84	20756	7.052	ug/l	98
37) Ethyl Acetate	4.714	43	60829	14.525	ug/l	98
43) Isopropyl Acetate	6.342	43	255611	36.029	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031723\
Data File : VX034605.D
Acq On : 18 Mar 2023 02:32
Operator : JC/MD
Sample : 01932-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CE-43-031423

Quant Time: Mar 20 02:07:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

